



TABLE 1. SUMMARY OF THE DATED SAMPLES.

Dr. A Ostlund
Ekelidsvagen, 5
S 45740, Fjallbacka
Sweden

21 October 2020

Rif.CEDAD: 2020_0148

Results of Radiocarbon Dating

Dear sir, please find enclosed the results of the radiocarbon dating of the samples you submitted to CEDAD (AMS and radiocarbon dating facility, University of Lecce, Italy) and listed in Table 1.

<i>Sample ID</i>	<i>CEDAD Code</i>	<i>Provenance</i>
2006P20	LTL20334A	
2006P27	LTL20335A	
2006P45	LTL20336A	
2006P51	LTL20337A	
2006P53	LTL20338A	
2006P85	LTL20339A	

Macro contaminants were removed from the samples by mechanical handpicking under optical microscope. The selected portion of the samples was treated in order to chemically remove any possible source of contamination.

The purified sample material was then converted to carbon dioxide by combustion in sealed quartz tubes. The obtained carbon dioxide was converted at 550°C into graphite by using ultrahigh purity Hydrogen as reducing medium and 2 mg iron powder as catalyst.



TABLE 2. MEASURED VALUES.

The sample yielded enough graphite to allow an accurate determination of the radiocarbon age by the accelerator mass spectrometer.

The radiocarbon concentrations have been determined in the accelerator mass spectrometer by comparing the ^{12}C , ^{13}C currents and the ^{14}C counts obtained from the samples with those obtained from standard materials supplied by IAEA (International Atomic Energy Agency) and NIST (National Institute of Standard and Technology).

The "conventional radiocarbon age" was calculated with a $\delta^{13}\text{C}$ correction based on the $^{13}\text{C}/^{12}\text{C}$ ratio measured directly with the accelerator. For the estimation of the measurement uncertainty (standard deviation) both the radioisotope counting statistics and the scattering of the data have been taken into account. The larger of the two is given as final error in Table 2.

Sample	Radiocarbon Age (BP)	$\delta^{13}\text{C}$ (‰)^(**)	Note
LTL20334A	3086 ± 45	-20.4 ± 0.5	
LTL20335A	2575 ± 45	-26.3 ± 0.6	
LTL20336A	2324 ± 45	-23.4 ± 0.4	
LTL20337A	2318 ± 45	-27.7 ± 0.3	
LTL20338A	1651 ± 45	-26.5 ± 0.4	
LTL20339A	3369 ± 45	-29.1 ± 0.5	

(**) The listed values of the carbon stable isotopes fractionation term ($\delta^{13}\text{C}$) are measured by AMS. These values can differ from the natural fractionation and from those measured by IRMS.

The conventional radiocarbon ages of the samples were converted into calendar years by using the software OxCal Ver. 3.5 based on the last atmospheric dataset [Reimer PJ, et al. 2013 *Radiocarbon* 55 No. 4-1869-1887]. The results of the calibration are reported in the following figures.



UNIVERSITÀ
DEL SALENTO

Dipartimento di Matematica e Fisica
"Ennio De Giorgi"



Centro di DATazione e Diagnostica




Figure 1. Calibration of the radiocarbon age of the sample
LTL20334A.




Figure 2. Calibration of the radiocarbon age of the sample LTL20335A.




Figure 3. Calibration of the radiocarbon age of the sample LTL20336A.




Figure 4. Calibration of the radiocarbon age of the sample LTL20337A.




Figure 5. Calibration of the radiocarbon age of the sample LTL20338A.




Figure 6. Calibration of the radiocarbon age of the sample LTL20339A.

Best Regards,

Prof. Dr. Lucio Calcagnile

Director, Centro di Datazione e Diagnostica dell'Università del Salento